



Republic of the Philippines
 Department of Agriculture
 Bureau of Fisheries and Aquatic Resources
BIDS AND AWARDS COMMITTEE OFFICE

2/F Fisheries Bldg. Complex, BPI Compound, Brgy. Vasra, Visayas Ave., Quezon City, Philippines
 Tel. No.: +02. 332. 4661 website: www.bfar.da.gov.ph e-mail: bac.eps@bfar.da.gov.ph

NOTICE OF AWARD

BIOKITS PHILIPPINES, INC.
 G/F OTM Building,
 No. 71 Scout Tuason Street,
 Brgy. South Triangle, Quezon City
 1103

Bid Reference No. 2022-27 entitled **"SUPPLY, DELIVERY, AND INSTALLATION OF MULTI-ANALYTE QUANTITATIVE DRUG RESIDUE ANALYZER FOR NFLD"**.

We are pleased to notify you that the contract of the above Project is hereby awarded to your firm as the **Single Calculated and Responsive Bidder (SCRB)** and compliant with the Technical Specifications required by the end-user at Contract Price specified in the attached summary.

You are therefore required, within **ten (10)** days from the receipt of this Notice of Award, to formally enter into a contract with us, and to submit the *Performance Security* in the form and the amount stipulated in the Instructions to Bidders.

Failure to enter into the said contract or provide the Performance Security shall constitute a sufficient ground for cancellation of this award and forfeiture of your Bid Security.

Very truly yours,


LILIBETH L. SIGNEY, PH.D., CESO V
 Officer-in-Charge, BFAR

I acknowledge receipt of this Notice of Award on the date indicated below:

Signature of Bidder's Authorized Representative: _____

Name of Bidder's Authorized Representative: _____

Date: _____

EBWIN T. NASOL
August 3, 2022



ANNEX TO NOTICE OF AWARD

SUMMARY OF ITEMS TO BE AWARDED TO:
BIOKITS PHILIPPINES INC.

BID REFERENCE NO.: 2022-27

“SUPPLY, DELIVERY, AND INSTALLATION OF MULTIAN-ALYTE QUANTITATIVE DRUG RESIDUE ANALYZER FOR NFLD”

| LOT NO. | DESCRIPTION | APPROVED BUDGET FOR THE CONTRACT (ABC) (PHP) | CONTRACT PRICE (PHP) |
|---------|---|--|----------------------|
| 1 | “SUPPLY, DELIVERY, AND INSTALLATION OF MULTI-ANALYTE QUANTITATIVE DRUG RESIDUE ANALYZER FOR NFLD” | ₱ 6,230,000.00 | ₱ 6,159,000.00 |

TECHNICAL SPECIFICATION

| No. | Item Description | Quantity | Unit |
|-----|---|----------|------|
| 1 | <p style="text-align: center;">Multi-Analyte Quantitative Drug Residue Analyzer with accessories and test kits</p> <p>A. Multi-Analyte Quatitative Drug Residue Analyzer</p> <p><u>Accreditation:</u> Internally accredited to full CE and UL certification</p> <p><u>Calibration method:</u> 9 point calibration</p> <p><u>Connectivity:</u> LIMS integration</p> <p><u>Data back-up methods:</u> Writable DVD, CD, USB Mass-storage or Network folder</p> <p><u>Measurement principle:</u> Chemiluminescent</p> <p><u>Network services:</u> Highly Secure Remote Diagnostics, automated software, and updates</p> <p><u>Peripherals:</u> Must be approved for use with the multi-residue analyzer</p> <p>Barcode scanner - for data input of scanned biochip arrays into the software</p> <p>Carrier handling tray - must be compatible with Antimicrobials and Anthelmintics,</p> <p><u>Biochip Arrays:</u> must be able to hold 9 individual biochips per carrier</p> | 1 | unit |

Thermoshaker: for biochip incubation;

- must provide optimum heating environment for samples (e.g. faster heat-up times, increased temperature range, and standardization of assay incubation conditions)

Sample loading: Single carrier loading tray

B. Test Kits:

1. Antimicrobial Array assays:

Quantity: 2

- Quinolones, Ceftriaxone, Thiamphenicol, Streptomycin, Tylosin, Tetracyclines
- 6 test analytes/sample, 45 samples/kit
- with certificate of analysis

2. Anthelmintics Array assays:

Quantity: 2

- Benzimidazoles, Amino Benzimidazoles, Thiabendazole, Triclabendazole, Levamisole, Moxidectin (MXD), Avermectins (AVM)
- 7 test analytes/sample, 45 samples/kit
- with certificate of analysis

C. PC and Imaging Software

1. Desktop computer

- 8th Gen Intel® Core™ i5-8500
- Windows 11 Pro
- Intel Integrated Graphics
- 16 GB Memory
- 256 GB SSD
- Optical Drive DVD +/-RW
- Bluetooth and Wi-Fi enabled
- 1 x USB 3.1 Gen 2 Type-C
- 5 x USB 3.1 Gen 1 Type-A
- 4 x USB 2.0 Type-A
- Monitor, keyboard, mouse

2. Printer

- **Functions:** Print, copy, scan, and fax
- Ink tank system

NOTES:

The Bidder must submit the following:

- Certificate of Accreditation to full CE (European Commission certification) and Underwriters Laboratories (UL);
- Certificate of Authorization to distribute from principal supplier;
- Certificate of Accreditation to ISO 13485 for test kits manufacturing.
- Test Kits must have at least **six (6)** months expiry upon delivery, must provide Certificate Analysis upon delivery and must conform to the criteria specified by the manufacturer.
- Bidder must provide preventive maintenance and calibration of the equipment up to **three (3)** years upon installation.
- Bidder must provide at least **two (2)** years warranty on the equipment, parts, and accessories.
- Bidder must conduct free training from manufacturer/principal supplier on the usage of equipment.
- Bidder must provide technical support on the conduct of method validation and accreditation of the method for PNS ISO/IEC 17025:2017 standard.
- Packaging Requirement: Packaging of materials or equipment upon arrival at the BFAR-National Fisheries Laboratory Division must conform to the criteria specified by the manufacturer.

| LOT NO. | DESCRIPTION | QUANTITY | DELIVERED, WEEKS/MONTHS | PLACE OF DELIVERY |
|---------|---|----------|---|---|
| 1 | "SUPPLY, DELIVERY, AND INSTALLATION OF MULTI-ANALYTE QUANTITATIVE DRUG RESIDUE ANALYZER FOR NFLD" | 1 Lot | within ninety (90) calendar days from receipt of the Notice to Proceed | BFAR-National Fisheries Laboratory Division (NFLD) – Residue Laboratory, 860 Arcadia Building, Quezon Avenue, Quezon City |

******NOTHING FOLLOWS******